

4.2 GENERAL EXPLANATION OF RULES/PRACTICAL ISSUES/EXAMPLES

Rule 1 - Devices that either do not touch the patient or contact intact skin only

General explanation of the rule

This is a fallback rule applying to all devices that are not covered by a more specific rule.

This is a rule that applies in general to devices that come into contact only with intact skin or that do not touch the patient.

| RULE 1 | EXAMPLES |
|--|---|
| <p>All non-invasive devices are in Class I, unless one of the rules set out hereinafter applies.</p> | <ul style="list-style-type: none"> - Body liquid collection devices intended to be used in such a way that a return flow is unlikely (e.g. to collect body wastes such as urine collection bottles, ostomy pouches, incontinence pads or collectors used with wound drainage devices). They may be connected to the patient by means of catheters and tubing. - Devices used to immobilize body parts and/or to apply force or compression on them (e.g. non-sterile dressings used to aid the healing of a sprain, plaster of Paris, cervical collars, gravity traction devices, compression hosiery). - Devices intended in general for external patient support (e.g. hospital beds, patient hoists, walking aids, wheelchairs, stretchers, dental patient chairs). - Corrective glasses, frames, stethoscopes for diagnosis, eye occlusion plasters, incision drapes, conductive gels, non-invasive electrodes (electrodes for EEG or ECG), image intensifying screens. - Permanent magnets for removal of ocular debris |

Practical issues of classification

Some non-invasive devices are indirectly in contact with the body and can influence internal physiological processes by storing, channeling or treating blood, other body liquids or liquids which are returned or infused into the body or by generating energy that is delivered to the body. These must be excluded from the application of this rule and be handled by another rule because of the hazards inherent in such indirect influence on the body.

Rule 2 - Channeling or storing for eventual administration

General explanation of the rule

These types of devices must be considered separately from the non-contact devices of rule 1 because they may be indirectly invasive. They channel or store substances that will be eventually delivered into the body. Typically these devices are used in transfusion, infusion, extracorporeal circulation, delivery of anaesthetic gases and oxygen.

In some cases devices covered under this rule are very simple gravity activated delivery devices.

| RULE 2 | EXAMPLES |
|--|---|
| <p>All non-invasive devices intended for channeling or storing blood, body liquids or tissues, liquids or gases for the purpose of eventual infusion, administration or introduction into the body are in Class IIa:</p> <p>- if they may be connected¹ to an active medical device in Class IIa or a higher class,</p> | <ul style="list-style-type: none"> - Devices intended to be used as channels in active drug delivery systems, e.g. tubing intended for use with an infusion pump. - Devices used for channeling, e.g. antistatic tubing for anesthesia, anesthesia breathing circuits and pressure indicator, pressure limiting devices. - Syringes for infusion pumps. |
| <p>- if they are intended for use for storing or channelling blood or other body liquids or for storing organs, parts of organs or body tissues (are in Class II a)</p> | <ul style="list-style-type: none"> - Devices intended to channel blood (e.g. in transfusion, extracorporeal circulation). - Devices intended for temporary storage and transport of organs for transplantation. - Devices intended for long term storage of biological substances and tissues such as corneas, sperm, human embryos, etc. |
| <p>in all other cases they are in Class I.</p> | <ul style="list-style-type: none"> - Devices that provide a simple channeling function, with gravity providing the force to transport the liquid, e.g. administration sets for infusion. - Devices intended to be used for a temporary containment or storage function such as cups and spoons specifically intended for administering medicines². - Syringes without needles |

Practical issues of classification

Blood bags are covered as an exception under a separate rule (see rule 18).

If a device, e.g. tubing, can be used for a purpose that would cause it to be connected to an active device such a device will be automatically in Class II A, unless the manufacturer clearly state that it should not be connected to an active device of Class II A or higher.

Explanation of special concepts

Note 1: "May be connected to an active device". Such connection is deemed to exist between a non-active device and an active device where a non-active device forms a link in the transfer of the substance between the patient and the active device and the safety and performance of one of the devices is influenced by the other device. For instance, this applies to tubing in an extracorporeal circulation system which is downstream from a blood pump and in the same blood flow circuit, but not directly in contact with the pump.

Note 2: Solutions intended for preservation of organs during storage and transport are not medical devices.

Rule 3 - Devices that modify biological or chemical composition of blood, body liquids or other liquids

General explanation of the rule

These types of devices must be considered separately from the non-contact devices of rule 1 because they are indirectly invasive. They treat or modify substances that will be eventually delivered into the body. This rule covers mostly the more sophisticated elements of extracorporeal circulation sets, dialysis systems and autotransfusion systems as well as devices for extracorporeal treatment of body fluids which may or may not be reintroduced immediately into the body, including, where the patient is not in a closed loop with the device..

| RULE 3 | EXAMPLES |
|---|---|
| All non-invasive devices intended for modifying the biological or chemical composition of blood, other body liquids or other liquids intended for infusion into the body are in Class Ib, | <ul style="list-style-type: none">- Devices intended to remove undesirable substances out of the blood by exchange of solutes such as hemodialyzers.- Devices intended to separate cells by physical means, e.g. gradient medium for sperm separation.- Haemodialysis concentrates. |
| unless the treatment consists of filtration, centrifugation or exchange of gas or heat, in which case they are in Class IIa. | <ul style="list-style-type: none">- Particulate filtration of blood in an extracorporeal circulation system. These are used to remove particles and emboli from the blood.- Centrifugation of blood to prepare it for transfusion or autotransfusion.- Removal of carbon dioxide from the blood and/or adding oxygen.- Warming or cooling the blood in an extracorporeal circulation system. |

Practical issues of classification

These devices are normally used in conjunction with an active medical device covered under rule 9 or rule 11.

Filtration and centrifugation should be understood in the context of this rule as exclusively mechanical methods.

Rule 4 - Devices in contact with injured skin

General explanation of the rule

This rule is intended to cover primarily wound dressings independently of the depth of the wound. The traditional types of products (e.g. used as a mechanical barrier) are well understood and do not result in any great hazard. There have also been rapid technological developments in this area, with the emergence of new types of wound dressings for which non-traditional claims are made, e.g. management of the micro-environment of a wound to enhance its natural healing mechanism. More ambitious claims relate to the mechanism of healing by secondary intent, such as influencing the underlying mechanisms of granulation or epithelial formation or preventing contraction of the wound. Some devices used on breached dermis may even have a life-sustaining or life-saving purpose, e.g. when there is full thickness destruction of the skin over a large area and/or systemic effect. Dressings containing medicinal products acting as ancillary to the dressing fall within Class III under Rule 13.

| RULE 4 | EXAMPLES |
|---|--|
| All non-invasive devices which come into contact with injured skin: - are in Class I if they are intended to be used as a mechanical barrier, for compression or for absorption of exudates, | - Wound dressings, such as absorbent pads, island dressings, cotton wool, wound strips and gauze dressings to act as a barrier or to maintain the wound positionally or to absorb exudates from the wound. |
| - are in Class IIb if they are intended to be used principally with wounds which have breached the dermis and can only heal by secondary intent | - Are principally intended to be used with severe wounds that have substantially and extensively breached the dermis, and where the healing process can only be by secondary intent such as: - dressings for chronic extensive ulcerated wounds - dressings for severe burns having breached the dermis and covering an extensive area - dressings for severe decubitus wounds - dressings incorporating means of augmenting tissue and providing a temporary skin substitute |
| - are in Class IIa in all other cases, including devices principally intended to manage the micro-environment of a wound. | - Have specific properties intended to assist the healing process by controlling the level of moisture at the wound during the healing process and to generally regulate the environment in terms of humidity and temperature, levels of oxygen and other gases and ph values or by influencing the process by other physical means . - These devices may specify particular additional healing properties whilst not being intended for extensive wounds requiring healing by secondary intent. - Adhesives for topical use. - Polymer film dressings, hydrogel dressings and non-medicated impregnated gauze dressings. |

Practical issues of classification

Products covered under this rule are extremely claim sensitive, e.g. a polymeric film dressing would be in Class II A if the intended use is to manage the micro-environment of the wound and in Class I if its intended use is limited to retaining an invasive cannula at the wound site. Consequently it is impossible to say a priori that a particular type of dressing is in a given class without knowing its intended use as defined by the manufacturer. However, a claim that the device is interactive or active with respect to the wound healing process usually implies that the device is in Class II B.

Most dressings that are intended for a use that is in Class II A or II B, also perform functions that are in Class I, e.g. that of a mechanical barrier. Such devices are nevertheless classed according to the intended use in the higher class.

For such devices incorporating medicines see rule 13 or animal tissues see rule 17.

Explanation of special concepts

- Breached dermis: the wound exposes at least partly the subcutaneous tissue.
- Secondary intent: the wound heals by first being filled with granulation tissue, subsequently the epithelium grows back over the granulation tissue and the wound contracts. In contrast primary intent implies that the edges of the wound are close enough or pulled together, e.g. by suturing, to allow the wound to heal.

Rule 5 Devices invasive in body orifices

General explanation of the rule

Invasiveness with respect to the body orifices (ear, mouth, nose, eye, anus, urethra and vagina) must be considered separately from invasiveness that penetrates through a cut in the body surfaces (surgical invasiveness). For short term use, a further distinction must be made between invasiveness with respect to the less vulnerable anterior parts of the ear, mouth and nose and the other anatomical sites that can be accessed through natural body orifices.

The surgically created stoma, which for example allows the evacuation of urine or faeces, should also be considered as a body orifice.

Devices covered by this rule tend to be diagnostic and therapeutic instruments used in particular specialities (ENT, ophthalmology, dentistry, proctology, urology and gynecology).

| RULE 5 | EXAMPLES |
|---|--|
| All invasive devices with respect to body orifices, other than surgically invasive devices and which are not intended for connection to an active medical device: | |
| - are in Class I if they are intended for transient use, | - Handheld mirrors used in dentistry to aid in dental diagnosis and surgery, dental impression materials, tubes used for pumping the stomach, impression tray, enema devices, examination gloves and prostatic balloon dilation catheters. |
| - are in Class IIa if they are intended for short term use, | - Contact lenses, urinary catheters, tracheal tubes, stents, vaginal pessaries and perineal reeducation devices. |
| except if they are used in the oral cavity as far as the pharynx, in an ear canal up to the ear drum or in a nasal cavity , in which case they are in Class I, | - Dressings for nose bleeds, dentures intended to be removed by the patient. |
| - are in Class IIb if they are intended for long term use, | - Urethral stents. |

| | |
|---|---|
| <p>except if they are used in the oral cavity as far as the pharynx, in an ear canal up to the ear drum or in a nasal cavity and are not liable to be absorbed by the mucous membrane, in which case they are in Class IIa.</p> | <p>- Orthodontic wire, fixed dental prostheses, fissures sealants.</p> |
| <p>All invasive devices with respect to body orifices, other than surgically invasive devices, intended for connection to an active medical device in Class IIa or a higher class, are in Class IIa.</p> | <p>- Tracheostomy or tracheal tubes connected to a ventilator, blood oxygen analysers placed under the eye-lid, powered nasal irrigators, nasopharyngeal airways, some enteral feeding tubes, fibreoptics in endoscopes connected to surgical lasers, suction catheters or tubes for stomach drainage, dental aspirator tips.</p> |

Rule 6 - Surgically invasive devices for transient use

General explanation of the rule

This rule covers principally three major groups of devices: devices that are used to create a conduit through the skin (needles, cannulae, etc.), surgical instruments (scalpels, saws, etc.) and various types of catheters, suckers, etc.

| RULE 6 | EXAMPLES |
|--|---|
| All surgically invasive ¹ devices intended for transient use are in Class IIa unless they are: | - Needles used for suturing, needles of syringes, lancets, suckers, single use scalpels, single use scalpel blades, support devices in ophthalmic surgery, staplers, surgical swabs, drill bits connected to active devices, surgical gloves, etchants, tester of artificial heart valves, heart valve occluder, heart valve sizers and holders, trial hip prosthesis heads or stems, swabs to sample exudates, single use aortic punches (see note 2) |
| - intended specifically to diagnose, monitor or correct a defect ² of the heart or of the central circulatory system ¹ through direct contact with these parts of the body, in which case they are in Class III ³ | - Cardiovascular catheters (e.g. angioplasty balloon catheters), including related guidewires and dedicated ⁴ disposable cardiovascular surgical instruments e.g. electrophysiological catheters, electrodes for electrophysiological diagnosis and ablation. - Catheters containing or incorporating sealed radioisotopes, where the radioactive isotope as such is not intended to be released into the body, if used in the central circulatory system |
| - reusable surgical instruments ¹ , in which case they are in Class I ³ , | - Scalpels, scalpel handles, drill bits, saws, that are not intended for connection to an active device, and retractors forceps, excavators and chisels. |
| - intended to supply energy in the form of ionizing radiation in which case they are in Class IIb, | - Catheters containing or incorporating sealed radioisotopes, where the radioactive isotope as such is not intended to be released into the body, if used in the circulatory system, excluding the central circulatory system |
| - intended to have a biological ⁵ effect or to be wholly or mainly absorbed ⁴ in which case they are in Class IIb, | |

| | |
|--|---|
| <p>- intended to administer medicines by means of a delivery system, if this is done in a manner that is potentially hazardous⁶ taking into account of the mode of application, in which case they are Class IIb.</p> | <p>- Devices for repeated self-application where dosage levels and the nature of the medicinal product are critical, e.g. insulin pens.</p> |
|--|---|

Explanations of special concepts

Note 1: Terms such as "surgically invasive device", "central circulatory system" and "reusable surgical instruments" are defined in Section I of Annex IX of the Directive. In particular surgical instruments connected to an active device are not considered to be "reusable surgical instruments".

Note 2: The expression "correct a defect" does not cover devices that are used accessorially in heart surgery, e.g. clamps. The first indent of this rule does not apply to aortic punches and similar cutting instruments which perform a similar function to a scalpel.

Note 3: Surgical instruments which are not specifically intended for purposes described in the first indent, and irrespective of the site of application, are in class IIA, if they are intended for single use and in class I if they are reusable.

Note 4: Dedicated means that the intended purpose of the device is to diagnose, monitor or correct a defect of the heart or of the central circulatory system.

Note 5:

Biological effect: All materials and devices have the potential to affect tissues following use in a surgically invasive procedure. A material is considered to have a biological effect if it actively and intentionally induces, alters or prevents a response from the tissues that is mediated by specific reactions at a molecular level. Such a device may be described as bioactive.

Wholly or mainly absorbed: The term absorption refers to the degradation of a material within the body and the metabolic elimination of the resulting degradation products from the body.

Note 6: The concept of "potentially hazardous manner" is related to the characteristics of the device and not the competence of the user.

Rule 7 - Surgically invasive devices for short-term use

General explanation of the rule

These are mostly devices used in the context of surgery or post-operative care (e.g. clamps, drains), infusion devices (cannulae, needles) and catheters of various types.

| RULE 7 | EXAMPLES |
|---|--|
| All surgically invasive devices intended for short term use are in Class IIa unless they are intended: | - Clamps, infusion cannulae, skin closure devices, temporary filling materials. - Tissue stabilisers ² used in cardiac surgery |
| - either specifically to diagnose, monitor or correct a defect ² of the heart or of the central circulatory system through direct contact with these parts of the body, in which case they are in Class III, | - Cardiovascular catheters, cardiac output probes and temporary pacemaker leads. - Thoracic catheters intended to drain the heart, including the pericardium - Carotid artery shunts |
| - or specifically for use in direct contact with the central nervous system, in which case they are in Class III, | - Neurological catheters, cortical electrodes and connonoid paddles. |
| - or to supply energy in the form of ionizing radiation in which case they are in Class IIb, | - Brachytherapy devices. |
| - intended to have a biological effect or to be wholly or mainly absorbed in which case they are in Class III, | - Absorbable sutures and biological adhesives. |
| - or to undergo chemical change in the body, except if the devices are placed in the teeth, or to administer medicines ¹ , in which case they are Class IIb. | - Adhesives |

Practical issues of classification

Note 1: Administration of medicines is more than just channeling, it implies also storage and/or influencing the volume and rate of the medicine delivered. Implanted capsules for the slow release of medicines are medicines and not medical devices.

Note 2: The expression “correct a defect” does not cover devices that are used accessorially in heart surgery, e.g. tissue stabilisers.

Rule 8 - Surgically invasive devices for long-term use and implantable devices

General explanation of the rule

These are mostly implants in the orthopaedic, dental, ophthalmic and cardiovascular fields as well as soft tissue implants such as implants used in plastic surgery.

| RULE 8 | EXAMPLES |
|---|--|
| All implantable devices and long-term surgically invasive devices are in Class IIb unless they are intended: | - Prosthetic joint replacements, ligaments, shunts, stents, nails, plates, intra-ocular lenses, internal closure devices, tissue augmentation implants, infusion ports, peripheral vascular grafts, penile implants, non-absorbable sutures, bone cements and maxillo-facial implants, visco-elastic surgical devices intended specifically for ophthalmic anterior segment surgery ¹ . |
| - to be placed in the teeth ² in which case they are in Class IIa, | - Bridges, crowns, dental filling materials and pins, dental alloys, ceramics and polymers. |
| - to be used in direct contact with the heart, the central circulatory system or the central nervous system, in which case they are Class III, | - Prosthetic heart valves, aneurysm clips, vascular prostheses, spinal stents, vascular stents, CNS electrodes and cardiovascular sutures. - Permanent vena cava filters |
| - to have a biological effect or to be wholly or mainly absorbed, in which case they are in Class III, | - Absorbable sutures, adhesives and implantable devices claimed to be bioactive through the attachment of surface coatings such as phosphorylcholine. |
| - or to undergo chemical change ³ in the body, except if the devices are placed in the teeth, or to administer medicines, in which case they are in Class III. | - Rechargeable non-active drug delivery systems. |

Practical issues of classification

Note 1 :these products are implants because in normal conditions a significant amount of the substance remains at the surgical site after the procedure. If these devices contain animal tissues or derivatives of animal tissues, they are covered by rule 17.

Note 2: Implants without bioactive coatings intended to secure teeth or prostheses to the maxillary or mandibular bones are in Class II B following the general rule. Hydroxy-apatite is considered as having biological effect only if so claimed and demonstrated by the manufacturer.

Note 3: The clause about chemical change under this rule does not apply to products such as bone cements where the chemical change takes place during the placement and does not continue in long term.

Rule 9 - Active therapeutic devices intended to administer or exchange energy

General explanation of the rule

Devices classified by this rule are mostly electrical equipment used in surgery such as lasers and surgical generators. In addition there are devices for specialised treatment such as radiation treatment. Another category consists of stimulation devices, although not all of them can be considered as delivering dangerous levels of energy considering the tissue involved.

| RULE 9 | EXAMPLES |
|---|---|
| All active therapeutic devices intended to administer or exchange energy are in Class IIa | <p><u>Electrical and/or magnetic and electromagnetic energy</u> - Muscle stimulators and external bone growth stimulators, TENS devices and eye electromagnets, electrical acupuncture</p> <p><u>Thermal energy</u> - Cryosurgery equipment, heat exchangers, except the types described below</p> <p><u>Mechanical energy</u> - Powered dermatomes, powered drills and dental hand pieces.</p> <p><u>Light</u> - Phototherapy for skin treatment and for neonatal care</p> <p><u>Sound</u> - Hearing aids</p> <p><u>Ultrasound</u> - Equipment for physiotherapy</p> |

| | |
|---|--|
| <p>unless their characteristics are such that they may administer or exchange energy to and from the human body in a potentially hazardous way¹, taking account of the nature, the density and the site of application of the energy, in which case they are in Class IIb.</p> | <p><u>Kinetic energy</u> - Lung ventilators</p> <p><u>Thermal energy</u> - Incubators for babies, warming blankets, blood warmers, electrically powered heat exchangers, for instance those used with patients incapable of reacting, communicating and/or who are without a sense of feeling</p> <p><u>Electrical energy</u> - High-frequency electrosurgical generators, and electrocautery equipment, including their electrodes, external pacemakers, external defibrillators, electroconvulsive therapy equipment</p> <p><u>Coherent light</u> - Surgical lasers</p> <p><u>Ultrasound</u> - Lithotriptors, surgical ultrasound devices</p> <p><u>Ionizing radiation</u> - Radioactive sources for afterloading therapy, therapeutic cyclotrons, linear accelerators, therapeutic X-ray sources.</p> |
| <p>All active devices intended to control and monitor the performance of active therapeutical devices in Class IIb or intended to influence directly the performance of such devices are in Class IIb.</p> | <p>- External feedback systems for active therapeutic devices, afterloading control devices.</p> |

Explanation of special concepts

Note 1: The decision as to whether a medical device administers or exchanges energy to and from the human body in a potentially hazardous way should take into account the following factors. The concept of "potentially hazardous" is dependent on the type of technology involved and the intended application of the device to the patient and not on the measures adopted by the manufacturer in view of good design management (e.g. use of technical standards, risk analysis). For instance all devices intended to emit ionizing radiation, all lung ventilators and lithotriptors should be in Class IIB. However, the manufacturer's obligation to comply with design requirements and solutions adopted, such as use of standards, exist independently from the classification system.

Rule 10 - Active devices for diagnosis.

General explanation of the rule

This covers principally a whole range of widely used equipment in the fields ultrasound diagnosis and capture of physiological signals as well as therapeutic and diagnostic radiology.

| RULE 10 | EXAMPLES |
|---|---|
| Active devices intended for diagnosis are in Class IIa: | |
| - if they are intended to supply energy which will be absorbed by the human body, except for devices used to illuminate the patient's body, in the visible spectrum, | - Magnetic resonance equipment, pulp testers, evoked response stimulators, diagnostic ultrasound. |
| - if they are intended to image in vivo distribution of radiopharmaceuticals, | - Gamma cameras, positron emission tomography and single photon emission computer tomography. |
| - if they are intended to allow direct diagnosis or monitoring of vital physiological processes ¹ , | - Electrocardiographs, electroencephalographs, cardioscopes with or without pacing pulse indicators ² - Electronic thermometers - Electronic stethoscopes - Electronic blood pressure measuring equipment |
| unless they are specifically intended for monitoring of vital physiological parameters, where the nature of variations is such that it could result in immediate danger to the patient, for instance variations in cardiac performance, respiration, activity of CNS in which case they are in Class IIb. | - Intensive care monitoring and alarm devices (for e.g. blood pressure, temperature, oxygen saturation), biological sensors, blood gas analysers used in open heart surgery, cardioscopes and apnea monitors, including apnea monitors in home care |
| Active devices intended to emit ionizing radiation and intended for diagnostic and therapeutic interventional radiology ³ including devices which control or monitor ⁴ such devices, or which directly influence their performance, are in Class II B. | - Diagnostic X-ray sources. |

Examples of special concepts:

Note 1: Vital physiological processes and parameters, include for example respiration, heart rate, cerebral functions, blood

gases, blood pressure and body temperature. Medical devices intended to be used for continuous surveillance of vital physiological processes in anesthesia, intensive care or emergency care are in Class IIB, whilst medical devices intended to be used to obtain readings of vital physiological signals in routine check ups and in self-monitoring are in Class IIA. A thermal imaging device intended to monitor blood flow is not considered to be a temperature measuring device.

Note 2: Devices specifically intended to monitor AIMDs fall under the AIMD Directive.

Note 3: Therapeutic interventional radiology refers to diagnosis being carried out during surgical procedures

Note 4: This refers to active devices for the control, monitoring or influencing of the emission of ionizing and not to the subsequent processing, recording or viewing of the resulting image.

Rule 11 - Active devices to administer, remove medicines and other substances to or from the body

General explanation of the rule

This rule is intended to cover primarily drug delivery systems and anesthesia equipment.

| <u>RULE 11</u> | <u>EXAMPLES</u> |
|--|--|
| All active devices intended to administer and/or remove medicines, body liquids or other substances to or from the body are in Class IIa, | - Suction equipment, feeding pumps. - Jet injectors for vaccination - Nebulisers to be used on conscious and spontaneously breathing patients where failure to deliver the appropriate dosage characteristics is not potentially hazardous |
| unless this is done in a manner: - that is potentially hazardous, taking account of the nature of the substances involved, of the part of the body concerned and of the mode of application, in which case they are in Class IIb. | - Infusion pumps, ventilators, anesthesia machines, anesthetic vaporisers, dialysis equipment, blood pumps for heart-lung machines, hyperbaric chambers, pressure regulators for medical gases, medical gas mixers, moisture exchangers in breathing circuits if used on unconscious or non-spontaneously breathing patients - Nebulisers where the failure to deliver the appropriate dosage characteristics could be hazardous. |

Rule 12 All other active devices

General explanation of the rule

This is a fallback rule to cover all active devices not covered by the previous rules.

| RULE 12 | EXAMPLES |
|--|---|
| All other active devices are in Class I. | <ul style="list-style-type: none">- Active diagnostic devices intended to illuminate the patient's body in the visible spectrum such as examination lights or to optically view the body such as surgical microscopes.- Devices intended in general for external patient support (e.g. hospital beds, patient hoists, walking aids, wheelchairs, stretchers, dental patient chairs).- Active diagnostic devices intended for thermography.- Active devices intended for recording, processing or viewing of diagnostic images.- Dental curing lights. |

4. Special rules

Rule 13 - Devices incorporating a medicinal substance (See MEDDEV. 2.1/3 rev. 1)

General explanation of the rule

This rule is intended to cover combination devices that contain a medicinal substance incorporated into the device for the purpose of assisting the functioning of that device. However this rule does not cover those devices incorporating substances which under other circumstances may be considered as medicinal substances, but which are incorporated into the device exclusively for the purpose at maintaining certain characteristics of the device and which are not liable to act on the body. For instance agents for the preservation of solutions for contact lenses (see MEDDEV 2.1/3 rev. 5.1 Sections A.3, 3.1 3rd note and b.1 9th paragraph). The primary function of the device does not rely on the pharmacological effect of the medicine. If the latter is the case, the product is a medicine rather than a device and not covered by this Directive.

| RULE 13 | EXAMPLES |
|---|--|
| All devices incorporating, as an integral part ¹ , a substance which, if used separately, can be considered to be a medicinal product as defined in Article 1 of the Directive 65/65/EEC, and which is liable to act on the human body with action ancillary to that of the devices, are in Class III. | <ul style="list-style-type: none">- Antibiotic bone cements, condoms with spermicide, heparin coated catheters, endodontic materials with antibiotics.- Ophthalmic irrigation solutions principally intended for irrigation, which contain components which support the metabolism of the endothelial cells of the cornea- Dressings incorporating an antimicrobial agent where the purpose of such an agent is to provide ancillary action on the wound |

Note 1: "Integral part" means that the device and the medicinal substance form one physical unit.

Rule 14 - Devices used for contraception or prevention of sexually transmitted diseases

General explanation of the rule

These intended uses relate to special cases of human vulnerability that cannot be covered by the normal criteria of time, invasiveness and organic function.

Although this rule covers two very different device applications, some devices may perform both functions, e.g. condoms. Devices intended to prevent the sexual transmission of the HIV are also covered by this rule.

| RULE 14 | EXAMPLES |
|---|--|
| All devices used for contraception or the prevention of the transmission of sexually transmitted diseases are in Class IIb, | - Condoms, contraceptive diaphragms. |
| unless they are implantable or long term invasive devices, in which case they are in Class III. | - Contraceptive intrauterine devices (IUDs) ¹ . |

Note 1: Intrauterine contraceptives whose primary purpose is to release progestogens are not medical devices.

Rule 15 - Specific disinfecting, cleaning and rinsing devices

General explanation of the rule

This rule is principally intended to cover various contact lens fluids. It also covers substances used principally in a medical environment to disinfect medical devices.

| RULE 15 | EXAMPLES |
|--|---|
| All devices intended specifically to be used for disinfecting, cleaning, rinsing or, when appropriate hydrating contact lenses are in Class IIb. | - Contact lens solutions, comfort solutions. |
| All devices intended specifically to be used for disinfecting medical devices are in Class IIa. | - Disinfectants specifically intended for instance for endoscopes or haemodialysis equipment, sterilizers specifically intended to sterilize medical devices in a medical environment and washer disinfectors. - Cleaners which disinfect prosthetic dentures. |
| This rule does not apply to products that are intended to clean medical devices other than contact lenses by means of physical action ¹ . | |

Practical issues of classification

Note 1: This rule does not apply to mechanical means of cleaning of devices, such as brushes and ultrasound. Such products will only fall under this directive if they are specifically intended for use with medical devices.

Rule 16 - Non-active devices to record X-ray diagnostic images

| RULE 16 | EXAMPLES |
|--|--|
| Non-active devices specifically intended for recording of X-ray diagnostic images are in Class II a. | X-ray films, photostimulable phosphor plates |

Note: This refers to primary recording media such as X-ray films and not to media used for subsequent reproduction.

Rule 17 - Devices utilizing animal tissues or derivatives

Explanation of the rule

This rule covers devices that contain or are made of animal tissues that have been rendered non-viable or derivatives from such tissues also being non-viable, i.e. where there is no longer any capacity for cellular metabolic activity. Devices containing non-inactivated animal tissues and/or any human tissues or derivatives are excluded from the scope of this Directive.

The manufacture of some devices may use industrial raw materials which contain small amounts of tallow or tallow derivatives (e.g. stearates in polymers). Such substances are not considered as derivatives of animal tissues for the purpose of this rule and therefore this rule does not apply.

| RULE 17 | EXAMPLES |
|---|--|
| All devices manufactured utilizing animal tissues or derivatives ¹ rendered non-viable are Class III except where such devices are intended to come into contact with intact skin ² only. | - Biological heart valves, porcine xenograft dressings, catgut sutures, implants and dressings made from collagen. |

Practical classification issues

- Devices made of non-viable animal tissue that come into contact with intact skin only (e.g. leather components of orthopedic appliances) are in Class I in accordance to rule 1.

Note 1: Derivatives are products that are processed from animal tissues and exclude substances such as milk, silk, beeswax, hair, lanolin

Note 2: Intact skin includes the skin around an established stoma unless the skin is breached.

Rule 18 - Blood bags

General explanation of the rule

This is a special rule that covers only blood bags.

| RULE 18 | EXAMPLES |
|--|--|
| By derogation from other rules, blood bags are in Class IIb. | - Blood bags (including those containing or coated with an anticoagulant). Where blood bags have a function greater than for storing purposes and include systems for preservation other than anti-coagulants then other rules (e.g. rule 13) may apply |

Note: Blood bags are described in the European Pharmacopoeia in the monograph on "Containers for Blood and Blood Components".